

Page 8 of 12

**WADSWORTH CENTER
NEW YORK STATE DEPARTMENT OF HEALTH
ENVIRONMENTAL LABORATORY APPROVAL PROGRAM**

Proficiency Test Report

Lab Id: 11642 USGS NATIONAL WATER QUALITY LABORATORY
POB 25046-MS 407-BLDG 95
EPA Lab Code: DENVER, CO-80225-0046
CO00009 UNITED STATES
(303) 236-2000
Director: MR. MERLE SHOCKEY

Shipment Date : 24-Jan-2005
Closing Date : 10-Mar-2005
Score Date : 29-Mar-2005

This report may contain data that are not covered by the NVLAP accreditation.

** Indicates NVLAP accredited analyte evaluated using the USEPA's National Standards for Water Proficiency Testing Studies Criteria Document.
NVLAP Lab Code 200387-0. ELAP is an A2LA accredited Proficiency Testing Provider. Certificate Number 1785.01

Shipment: 280 Non Potable Water Chemistry

Analyte Name	Units	Sample ID	Method	Result	Mean/Target *	Warning Limits	Acceptance Limits	Score
Sample: Non Potable Water Base/Neutrals								
Di-n-octyl phthalate EPA Code: N/A	ug/L	8036	USGS O-3116-87	72.6	84.3		15.7 - 153	Satisfactory 79 passed out of 79 reported results.
Fluoranthene EPA Code: N/A	ug/L	8036	USGS O-3116-87	166	143		69.9 - 216	Satisfactory 84 passed out of 87 reported results.
Hexachlorobenzene EPA Code: N/A	ug/L	8036	USGS O-3116-87	54.8	53.0		26.8 - 79.3	Satisfactory 78 passed out of 81 reported results.
Indeno(1,2,3-c,d)pyrene EPA Code: N/A	ug/L	8036	USGS O-3116-87	80.2	137		3.84 - 271	Satisfactory 85 passed out of 86 reported results.
Isophorone EPA Code: N/A	ug/L	8036	USGS O-3116-87	< 2.2	< 30.0 *			Satisfactory 74 passed out of 75 reported results.
Naphthalene EPA Code: N/A	ug/L	8036	USGS O-3116-87	107	96.0		30.5 - 162	Satisfactory 85 passed out of 85 reported results.
Nitrobenzene EPA Code: N/A	ug/L	8036	USGS O-3116-87	< 1.4	< 10.0 *			Satisfactory 74 passed out of 75 reported results.
Phenanthrene EPA Code: N/A	ug/L	8036	USGS O-3116-87	114	101		42.7 - 159	Satisfactory 87 passed out of 87 reported results.
Pyrene EPA Code: N/A	ug/L	8036	USGS O-3116-87	97.0	87.1		34.2 - 140	Satisfactory 86 passed out of 87 reported results.
Sample: Non Potable Water Total Cyanide								
Cyanide, Total ** EPA Code: N/A	mg/L	8011	USGS I-4302-85	0.368	0.359 *	0.282 - 0.429	0.245 - 0.465	Satisfactory 82 passed out of 89 reported results.
Sample: Non Potable Water Oil and Grease								
Oil & Grease Total Recoverable ** EPA Code: 0104	mg/L	8010	EPA 1664A	19.7	21.3 *	14.8 - 23.9	12.5 - 26.2	Satisfactory 90 passed out of 101 reported results.